UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/523,739	02/04/2005	Shigeru Nemoto	KITO2.006APC	7693
20995 7590 03/28/2007 KNOBBE MARTENS OLSON & BEAR LLP				INER
2040 MAIN STREET FOURTEENTH FLOOR IRVINE, CA 92614			STIGELL, THEODORE J	
			ART UNIT	PAPER NUMBER
			3763	
SHORTENED STATUTORY P	PERIOD OF RESPONSE	NOTIFICATION DATE	DELIVER	Y MODE
3 MONT	HS	03/28/2007	ELECTRONIC	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Notice of this Office communication was sent electronically on the above-indicated "Notification Date" and has a shortened statutory period for reply of 3 MONTHS from 03/28/2007.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jcartee@kmob.com eOAPilot@kmob.com

	Application No.	Applicant(s)				
	10/523,739	NEMOTO, SHIGERU				
Office Action Summary	Examiner	Art Unit				
	Theodore J. Stigell	3763				
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply						
A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication. - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).						
Status						
1) Responsive to communication(s) filed on 22 D	ecember 2006					
	action is non-final.					
	<u> </u>					
closed in accordance with the practice under <i>Ex parte Quayle</i> , 1935 C.D. 11, 453 O.G. 213.						
Disposition of Claims						
4)⊠ Claim(s) <u>1-12 and 14</u> is/are pending in the application.						
4a) Of the above claim(s) is/are withdrawn from consideration.						
5) Claim(s) is/are allowed.						
6)⊠ Claim(s) <u>1-12,14</u> is/are rejected.						
7) Claim(s) is/are objected to.						
8) Claim(s) are subject to restriction and/o	or election requirement					
Application Papers						
9)☐ The specification is objected to by the Examiner.						
10)☐ The drawing(s) filed on is/are: a)☐ accepted or b)☐ objected to by the Examiner.						
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).						
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).						
11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.						
Priority under 35 U.S.C. § 119						
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 						
Attachment(s)						
1) Notice of References Cited (PTO-892) 4) Interview Summary (PTO-413) Paper No(s)/Mail Date						
Notice of Draftsperson's Patent Drawing Review (PTO-948) Paper No(s)/Mail Date						

Application/Control Number: 10/523,739

Art Unit: 3763

DETAILED ACTION

Response to Amendment

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

Claims 1-12 and 14 are rejected under 35 U.S.C. 102(a) as being anticipated by Yamamoto (WO 03/039644). Yamamoto discloses a liquid injector for causing a syringe (1000) to suck a liquid from a liquid tank (2000) and inject the liquid into a patient, the syringe having a cylinder member and a piston member (1002) inserted slidably into the cylinder member comprising a patient tube (3010) having a leading end connected to a patient, a syringe tube (3001) connected to the syringe, a tank tube (3004) having a trailing end connected to the liquid tank, a tube connecting means (not numbered) for connecting a trailing end of the patient tube, a leading end of the syringe tube, and a leading end of the tank tube, a syringe drive mechanism for relatively moving the cylinder member and the piston member to cause the syringe to suck and inject fluid, a connection switch mechanism (101,102) for switching between a suck state in which the patient tube is blocked and the syringe tube is connected to the tank tube and an injection state in which the tank tube is blocked and the syringe tube is connected to the patient tube, an interlock control means for interlocking operation of the syringe drive mechanism and operation of the connection switch mechanism and an attachment tube

Page 3

(3008) connected to the patient tube, and an attachment block mechanism (103) for pressing and blocking the attachment tube to freely open or close the attachment tube, wherein the control means causes the attachment block mechanism to release the blocking only when the patient tube is blocked (Figure 13), wherein the connection switch mechanism can include a switching valve as is well-known in the art, wherein the connection switch mechanism has an injection block mechanism (103) and a suck block mechanism (101), wherein the connection switch mechanism also has an open or close mechanism (222,218) for interlocking open and close operation of the injection block mechanism, wherein the interlock control means causes the injection block mechanism to block the patient tube when the injection is completed, and further comprising an injection block sensor and a suck block sensor.

Response to Arguments

Applicant's arguments filed 12/22/2006 have been fully considered but they are not persuasive.

Applicant cannot rely upon the foreign priority papers to overcome this rejection because a translation of said papers has not been made of record in accordance with 37 CFR 1.55. See MPEP § 201.15.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within

Application/Control Number: 10/523,739

Art Unit: 3763

TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Theodore J. Stigell whose telephone number is 571-272-8759. The examiner can normally be reached on M-F 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Application/Control Number: 10/523,739

Art Unit: 3763

Page 5

Theodore J. Stigell

Mul